

# DECLARATION of CONFORMITY

We, NIKON INSTRUMENTS EUROPE B.V.  
TRIPOLIS 100, BURGERWEESHUISPAD 101, 1076 ER AMSTERDAM,  
THE NETHERLANDS ,

declare under our sole responsibility that the product,

TYPE of PRODUCT: Microscope with AC Adapter  
MODEL NUMBER: ECLIPSE Si RS  
with ATS018T-P050

to which this declaration relates is in conformity  
with the following standards or other normative documents,

EN 61010-1:2001, EN 61010-2-101:2002, EN 62471:2008

EN 62366:2008+A1:2015, EN 62304:2006

EN 61326-1:2013, EN 61326-2-6:2013

EN 55011:2009+A1:2010 Group 1, Class B

EN 61000-4-2:2009, EN 61000-4-3:2006+A1:2008+A2:2010

EN 61000-4-4:2012, EN 61000-4-5:2014+A1:2017

EN 61000-4-6:2014, EN 61000-4-8:2010

EN 61000-4-11:2004+A1:2017

EN 61000-3-2:2019, EN 61000-3-3:2013+A1:2019

EN 50581:2012

following the provisions of

IN VITRO DIAGNOSTIC MEDICAL DEVICE Directive (98/79/EC)

LOW VOLTAGE Directive (2014/35/EU)

EMC Directive (2014/30/EU)

RoHS Directive (2011/65/EU) .

Manufacturer's name:

NIKON CORPORATION  
HEALTHCARE BUSINESS UNIT

Location of the  
manufacturer:

471 NAGAODAI-CHO, SAKAE-KU  
YOKOHAMA-SHI, KANAGAWA  
244-8533 JAPAN

AMSTERDAM /

28 January, 2021



T. SASOKA, President