

DECLARATION of CONFORMITY

We, NIKON EUROPE B.V.
TRIPOLIS 100, BURGERWEESHUISPAD 101, 1076 ER AMSTERDAM,
THE NETHERLANDS,

declare under our sole responsibility that the product,
TYPE of PRODUCT: Stereo Microscope and LED Diascopic Illumination
Stand with AC Adapter: 3-AC, 4-AC
MODEL NUMBER: SMZ1270, SMZ800N and P-DSL32 with
FSP040-RHAN2, FSP040-RHAN3

to which this declaration relates is in conformity
with the following standards or other normative documents,

EN 61010-1:2010 EN 61010-2-101:2002

EN 62366:2008

EN 62471:2008

EN 61326-1:2013 , EN 61326-2-6:2013

EN 55011:2009+A1:2010 Group1, Class B

EN 61000-4-2:2009

EN 61000-4-3:2006+A1:2008+A2:2010

EN 61000-4-4:2012

EN 61000-4-5:2006

EN 61000-4-6:2009

EN 61000-4-8:2010

EN 61000-4-11:2004

EN 61000-3-2:2006+A1:2009+A2:2009

EN 61000-3-3:2008


EN 50581:2012

following the provisions of
IN VITRO DIAGNOSTIC MEDICAL DEVICE Directive (98/79/EC)
LOW VOLTAGE Directive (2014/35/EU)
EMC Directive (2014/30/EU)
RoHS Directive (2011/65/EU, 2015/863/EU).

Manufacturer's name: NIKON CORPORATION

Location of the
manufacturer: 471, Nagaodai-cho, Sakae-ku,
Yokohama, Kanagawa
244-8533 Japan

AMSTERDAM/
21 July, 2021


TAKAHARU SASOKA
Director & Executive Vice President
Healthcare Division Head