

DECLARATION of CONFORMITY

We, NIKON EUROPE B.V.
TRIPOLIS 100, BURGERWEESHUISPAD 101, 1076 ER AMSTERDAM,
THE NETHERLANDS,

declare under our sole responsibility that the product,
TYPE of PRODUCT: Inverted Research Microscope with AC Adapter:
3-AC, 4-AC
MODEL NUMBER: ECLIPSE Ti2-U with FSP040-RHAN2,
FSP040-RHAN3

to which this declaration relates is in conformity
with the following standards or other normative documents,


EN 61010-1:2001 EN 61010-2-101:2002
EN 62366:2008+A1:2015
EN 62471:2008
EN 61326-1:2013 , EN 61326-2-6:2013
EN 55011:2009+A1:2010 Group1, Class B
EN 61000-4-2:2009
EN 61000-4-3:2006+A1:2008+A2:2010
EN 61000-4-4:2012
EN 61000-4-5:2006
EN 61000-4-6:2014
EN 61000-4-8:2010
EN 61000-4-11:2004
EN 61000-3-2:2014
EN 61000-3-3:2013
EN 50581:2012

following the provisions of
IN VITRO DIAGNOSTIC MEDICAL DEVICE Directive (98/79/EC)
LOW VOLTAGE Directive (2014/35/EU)
EMC Directive (2014/30/EU)
RoHS Directive (2011/65/EU, 2015/863/EU).

Manufacturer's name: NIKON CORPORATION

Location of the
manufacturer: 471, Nagaodai-cho, Sakae-ku,
Yokohama, Kanagawa
244-8533 Japan

AMSTERDAM/
21 July, 2021


TAKAHARU SASAOKA
Director & Executive Vice President
Healthcare Division Head